

## Can a Paper-Based Device Trace COVID-19 Sources with Wastewater-Based Epidemiology?

Kang Mao, Hua Zhang,\* and Zhugen Yang\*



Cite This: *Environ. Sci. Technol.* 2020, 54, 3733–3735



Read Online

ACCESS |

Metrics & More

Article Recommendations

SCIENTIFIC  
OPINION  
NON-PEER  
REVIEWED



A recent outbreak of novel coronavirus pneumonia (COVID-19) caused by SARS-CoV-2 infection has spread rapidly around the globe, with cases now confirmed in 130 countries worldwide. Although public health authorities are racing to contain the spread of COVID-19 around the world, the situation is still grim. About 158 111 confirmed cases and 5946 cumulative deaths (81 059 confirmed cases and 3204 cumulative deaths from China) have been reported around the globe as of March 15, 2020. Some clinical cases have found that some carriers of the virus may be asymptomatic, with no fever, and no, or only slight symptoms of infection. Without the ability to screen these asymptomatic patients quickly and effectively, these unsuspecting carriers have the potential to increase the risk of disease transmission if no early effective quarantine measures are implemented. Therefore, to trace unknown COVID-19 sources, fast and accurate screening of potential virus carriers and diagnosis of asymptomatic patients is a crucial step for intervention and prevention at the early stage.

It remains a highly challenging logistical exercise for medical professionals to practically and effectively screen suspected infectious cases from individual households. Such a massive undertaking is time-consuming and labor intensive and is

constrained by the availability of testing technologies at this extremely critical time. However, an alternative method utilizing wastewater-based epidemiology (WBE), may provide an effective approach to predict the potential spread of the infection by testing for infectious agents in wastewater, which has been approved as an effective way to trace illicit drugs, and obtain information on health, disease, and pathogens.<sup>1</sup>

Faeces and urine from disease carriers in the community will contain many biomarkers that can enter the sewer system. A recent study demonstrated that live SARS-CoV-2 was isolated from the faeces and urine of infected people,<sup>2</sup> which would then enter the wastewater treatment system. A further study has shown that SARS-CoV-2 can typically survive for up to several days in an appropriate environment after exiting the human body. There is potential, therefore, that the analysis of SARS-CoV-2 in community wastewater could trace COVID-19 sources through sewage pipe networks and determine whether there are potential SARS-CoV-2 carriers in certain local areas. If SARS-CoV-2 can be monitored in the community at the early stage through WBE, effective intervention can be taken as early as possible to restrict the movements of that local population, working to minimize the pathogen spread and threat to public health.

Using a WBE approach in developing an early warning system and consequent effective intervention system will require a rapid analytical method for the on-site detection of viruses at the wastewater collection point. Currently, the most direct method for the detection of SARS-CoV-2 is a nucleic acid-based polymerase chain reaction (PCR) assay, which is also a means for confirmation of COVID-19 patients throughout China. Although PCR has high sensitivity and specificity, requirements for complicated sample handling in the laboratory, skilled personnel, and a long period of data processing and analysis (4–6 h) are not conducive to real-time and effective monitoring of samples on location. Therefore, it is critical to develop efficient transportable and robust analytical tools to accurately and quickly trace low-level

Received: February 25, 2020

Published: March 23, 2020



Table 1. Examples of Paper-Based Devices for Infectious Diseases and Pathogens Determination

infectious diseases/pathogens	characteristics of paper-based devices	detection method
malaria	paper device combined vertical flow sample-processing steps	visual UV/lateral flow device
rotavirus A	integrated nucleic acid test on a single paper device, including extraction, amplification, and on-site detection	naked eye
Zika virus	wax-printed paper devices utilizing isothermal amplification	smartphone
human papillomavirus	paper device in a foldable system allowing for fully integrated operation from sample to result	lateral flow device
HIV	paper devices fabricated with cellulose paper and flexible plastic plate	electrochemistry
<i>Neisseria meningitidis</i>	versatile paper devices integrated with isothermal amplification	visual fluorescence
<i>Listeria monocytogenes</i>	loop-mediated isothermal amplification (LAMP)-based paper devices	visual fluorescence
<i>Cochlodinium polykrikoides</i>	paper devices based on LAMP	visual fluorescence
<i>Staphylococcus aureus</i>	self-priming paper devices	visual fluorescence
<i>Vibrio parahaemolyticus</i>	self-priming paper devices	visual fluorescence
<i>Mycobacterium smegmatis</i>	paper devices combined thermal lysis and isothermal amplification into a single step	visual fluorescence
<i>Bacillus subtilis</i>	a wax-printed cellulose paper device	colorimetry
<i>Salmonella</i>	paper devices integrated with purification, amplification, and on-site detection	colorimetry
<i>Escherichia coli</i>	foldable paper devices with the ability of long-term reagents storage	colorimetry
	paper devices based on isothermal amplification and on-chip detection	visual fluorescence
	paper machine integrated sample preparation and isothermal amplification with end point detection	visual UV/camera
	paper devices integrated extraction, purification, amplification and detection	smartphone/naked eye
	paper devices combined thermal lysis and isothermal amplification	visual fluorescence
bovine infectious reproductive diseases	multiplexed and point-of-care paper-analytical device	visual UV/smartphone
highly pathogenic strain of porcine reproductive and respiratory syndrome virus (HP-PRRSV)	paper devices fabricated with filter paper and plastic chip	colorimetry

SARS-CoV-2 sources through WBE to confirm these suspected cases and screen asymptomatic infected cases without centralized laboratories.

Paper analytical devices have emerged as powerful tools for the rapid diagnosis of pathogens and determination of infection transmission.<sup>3</sup> The paper-based device is a small analytical tool with different functional areas printed with a wax printer that integrates all processes (extraction, enrichment, purification, elution, amplification, and visual detection) required for nucleic acid testing into an inexpensive paper material. The whole testing process can be completed through simple folding of a paper-based device in different ways in different steps without a pump or power supply, which overcomes the limitation of PCR and avoids multiple processes. Paper analytical devices enable multiplexed, sensitive assays that rival PCR laboratory assays and provide high-quality, fast precision diagnostics for pathogens. For example, a recent work has demonstrated that the multiplexed determination of malaria from whole blood using a paper-based device in rural Uganda.<sup>4</sup> The test could sensitively analyze multiplexed nucleic acid sequences of pathogens within 50 min, which gave a higher-quality and faster precision diagnosis for malaria than PCR.

In addition, paper analytical devices are easy to stack, store, and transport because they are thin, lightweight, and of different thicknesses. Visual analysis is made simple due to the strong contrast with a colored substrate. Paper-based devices can also be incinerated after use, reducing the risk of further contamination.

Although wastewater is a complex matrix, paper-based devices have shown the potential to detect pathogens in wastewater. We have developed a fast “sample-to-answer”

analysis method that can provide quantitative monitoring of nucleic acids and genetic information through the analysis of sewage,<sup>5</sup> which was confirmed with a robust electrophoresis and agarose gel image assay, showing promising reliability for wastewater analysis. Additional paper-based devices have also been fabricated for infectious diseases and pathogens determination as shown in Table 1.

In summary, the paper-based device has the potential to be used as a small, portable device to detect SARS-CoV-2 in wastewater on site and to track virus carriers in the community. Such an approach could provide near real-time and continuous data and serve as an early warning sensing system to help local governments and agencies make effective interventions to isolate potential virus carriers and prevent the spread of epidemics. We believe that in the case of asymptomatic infections in the community or people are not sure whether they are infected or not, rapid and real-time community sewage detection through paper analytical devices can determine whether there are SARS-CoV-2 carriers in the area in a timely manner to enable rapid screening, quarantine, and prevention. The potentially infected patient will also benefit from paper analytical device tracing SARS-CoV-2 sources with WBE, providing information for the correct and timely treatment of COVID-19.

## ■ AUTHOR INFORMATION

### Corresponding Authors

Hua Zhang – State Key Laboratory of Environmental Geochemistry, Institute of Geochemistry, Chinese Academy of Sciences, Guiyang 550081, China; CAS Center for Excellence in Quaternary Science and Global Change, Xi'an 710061, Shanxi Province, China; Email: zhanghua@mail.gyig.ac.cn

**Zhugen Yang** – School of Water, Energy and Environment,  
Cranfield University, Cranfield MK43 0AL, United Kingdom;  
orcid.org/0000-0003-4183-8160; Email: zhugen.yang@  
cranfield.ac.uk

#### Author

**Kang Mao** – State Key Laboratory of Environmental  
Geochemistry, Institute of Geochemistry, Chinese Academy of  
Sciences, Guiyang 550081, China; orcid.org/0000-0002-  
1777-5959

Complete contact information is available at:  
<https://pubs.acs.org/10.1021/acs.est.0c01174>

#### Notes

The authors declare no competing financial interest.

#### ACKNOWLEDGMENTS

This work is supported by STS of CAS (KFJ-STIS-QYZD-185) and Royal Academy of Engineering Frontier Follow-up Grant (FF\1920\1\36). Z.Y. thanks UK NERC Fellowship grant (NE/R013349/2). We thank the editors from *ES&T* for their help in manuscript revision.

#### REFERENCES

- (1) Yang, Z.; Kasprzyk-Hordern, B.; Frost, C. G.; Estrela, P.; Thomas, K. V. Community Sewage Sensors for Monitoring Public Health. *Environ. Sci. Technol.* **2015**, *49* (10), 5845–5846.
- (2) Holshue, M. L.; DeBolt, C.; Lindquist, S.; Lofy, K. H.; Wiesman, J.; Bruce, H.; Spitters, C.; Ericson, K.; Wilkerson, S.; Tural, A.; Diaz, G.; Cohn, A.; Fox, L.; Patel, A.; Gerber, S. I.; Kim, L.; Tong, S.; Lu, X.; Lindstrom, S.; Pallansch, M. A.; Weldon, W. C.; Biggs, H. M.; Uyeki, T. M.; Pillai, S. K. First Case of 2019 Novel Coronavirus in the United States. *N. Engl. J. Med.* **2020**, *382* (10), 929–936.
- (3) Magro, L.; Escadafal, C.; Garneret, P.; Jacquelin, B.; Kwasiborski, A.; Manuguerra, J. C.; Monti, F.; Sakuntabhai, A.; Vanhomwegen, J.; Lafaye, P.; Tabeling, P. Paper microfluidics for nucleic acid amplification testing (NAAT) of infectious diseases. *Lab Chip* **2017**, *17* (14), 2347–2371.
- (4) Reboud, J.; Xu, G.; Garrett, A.; Adriko, M.; Yang, Z.; Tukahebwa, E. M.; Rowell, C.; Cooper, J. M. Paper-based microfluidics for DNA diagnostics of malaria in low resource underserved rural communities. *Proc. Natl. Acad. Sci. U. S. A.* **2019**, *116* (11), 4834–4842.
- (5) Yang, Z.; Xu, G.; Reboud, J.; Kasprzyk-Hordern, B.; Cooper, J. M. Monitoring Genetic Population Biomarkers for Wastewater-Based Epidemiology. *Anal. Chem.* **2017**, *89* (18), 9941–9945.